Daily Measurements

Take one measurement each day unless advised otherwise by your healthcare provider. The LED will turn RED each morning and then turn GREEN once a measurement has been successfully transmitted.





Measurement Needed

Measurement Successful

For further assistance, please visit Tenovi.com/support.

tenoví

© 2025 Tenovi Co.

All rights reserved. Republication, duplication or modification in whole or in part is prohibited without prior written permission from Tenovi. All trademarks are owned by or licensed to Tenovi, Co. unless otherwise noted.

Power Your Gateway


Plug your Gateway into an outlet with the provided power cord and wait until LED on the Gateway turns red.




Insert BPM Batteries

Insert the four batteries into the back of your blood pressure monitor and connect the cuff into the port on the left back side of the monitor.

Battery Cover






2

Pairing Your Device

Press and hold the bluetooth button. The letter "P" will appear blinking. When the Gateway beeps and 4 squares are shown, your device has been successfully paired. This only needs to be done one time.






Body Position

Sit comfortably with your tested arm resting on a flat surface at the same level as your heart. Relax your arm and hand onto the table and turn your palm upwards.



Proper Set Up

Fasten the cuff to your left arm directly on your skin. Leave 1 inch between the cuff and your elbow joint. The hose should rest over your forearm



Take A Measurement

Press the "Start/Stop" button to take a measurement. If the LED ring is Green, you have properly taken a measurement for the day and that measurement has been successfully transmitted.



