

Pillbox

User's Guide



Important Safety and Product Information:



Warnings:

Failure to heed the warning below could result in an accident or medical event resulting in death or serious injury

This device contains electrical equipment that could cause injury if not handled properly.

Always consult your physician before beginning or modifying any healthcare program.

Do not allow children or pets to play with the device. The product contains small components that can be a choking hazard.

Do not attempt to replace any parts or open the enclosure. Doing so will void the warranty.

Do not use abrasive cleaners to clean your device.

Only use AA batteries to power the pillbox.

FCC Statement

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:
(1) This device may not cause harmful interference,

and (2) This device must accept any interference received, including interference that may cause undesired operation.

Warning: Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications.

However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by removing the batteries from the equipment.

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FCC Radiation Exposure Statement

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment



FCC ID: 2AZT4-TENPB001

FCC Radiation Exposure Statement


This device complies with Industry Canada licence-exempt RSS standard(s).

Operation is subject to the following two conditions:
(1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of the device."

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

(1) l'appareil ne doit pas produire de brouillage, et (2) l'utilisateur de l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement."

Getting To Know Your Device



Getting Started

Step 1: Plug in Gateway

- A** **Connect** the Gateway power cord to an electrical outlet.
- B** **Plug** the Gateway power cord into the back of the Gateway.

C

The Gateway LED will first flash purple for up to ten minutes. It will then turn red indicating that it is successfully connected to a cellular network.

Step 2: Connect your pillbox

Remove the pull tab from the battery compartment. Then **push the black button** on the bottom. The gateway should beep if completed properly. You only need to perform this step the first time you use your pillbox.



Warnings:

Ensure that the space between the pillbox and the Gateway is free from obstructions and that the pillbox and Gateway are in the same room when pairing.

Step 3: Refill your pillbox

Press the Refill button at the top of the pillbox to enter Refill Mode. You will see all of the pill compartments illuminate. **Fill your pillbox** by pressing down on each AM and PM compartment. Place medication taken in the morning in the daily AM compartments and medication taken in the evening in the daily PM compartments. To exit Refill Mode, **press the Refill button** again. You must enter Refill mode each week when you refill your pillbox with medication.

Step 4: Take your medication

To **take your daily medication** as directed by your doctor, press down on the respective AM or PM compartment. When you open any compartment, the pill compartment corresponding to the correct day and time will illuminate.

Step 5: Offload pillbox data

The pillbox will automatically detect when you open a compartment. This information will be saved by your pillbox. Whenever you are within 30 feet of the Tenovi Gateway, the saved information from your pillbox will automatically be sent to your care provider. At this time, the cellular Gateway will beep and turn green indicating that data has been sent properly.

Step 6: Use your pillbox daily

The Tenovi Pillbox will save all open events, even if you are away from the Tenovi Gateway. Whenever you open a pill compartment within 30 feet of the Tenovi Gateway, all stored open events will be sent to your care provider through the Tenovi Gateway. Your Gateway will turn from red to green and will beep if a medication event was sent that day.



Tenovi Health
1 Cate Street, Ste 100
Portsmouth, NH 03801
Email Contact: support@tenovi.com